

TITLE

CEOP_Tsukuba_TEREC-UT_20080701_20081231.flx

CONTACT

Jun Asanuma
Terrestrial Environment Research Center, University of Tsukuba
Tenno-dai 1-1-1
Tsukuba, Ibaraki 305-8572 Japan
E-mail: asanuma@suii.tsukuba.ac.jp

DATE OF THIS DOCUMENT

10 Apr. 2010

1. 0 DATASET OVERVIEW

1.1 Introduction

Terrestrial Environment Research Center (TEREC) started its long-term measurement of the surface heat and water balance at its surface heat and water balance field in 1981, and is continuing the observation. The field is circle in shape and has a 30-meter tower at its center. On the tower, profiles of air temperature, humidity and turbulence fluxes are measured and while, on the ground of the field, various measurements relevant to the surface heat and water budget are also being made. These include radiation components, ground water levels, soil temperature and soil heat flux. The measured data averaged over various time scale (30min., 1hour, 1day) are available on the web page (<http://www.suii.tsukuba.ac.jp>). Some of the can be dated back to 1980's. Among the measurements, basic meteorological and hydrological parameters are provided for CEOP2.

1.2 Time period covered by the data

Start: 1 July 2008, 00:00 UTC
End: 31 December 2008, 00:00 UTC

1.3 Temporal characteristics of the data

Original data, from which the data submitted to CEOP2, is averaged for 30 minutes.

1.4 Physical location of the measurement

Latitude: 36°06'35" N
Longitude: 140°06'00" E
Altitude: 27.0 m a.s.l.

1.5 Data source

Original data provided by the Terrestrial Environment Research Center, University of Tsukuba.

1.6 WWW address references

<http://www.suiri.tsukuba.ac.jp/>

2.0 INSTRUMENTATION DESCRIPTION

2.1 Platform

On the tower, thermometers are mounted at the 3 vertical levels, namely, 1.6 and 29.5 m above the ground. Also, 3 sonic anemometer-thermometers are mounted at 2 vertical levels, 1 at 1.6 m and 2 at 29.5m. The one of the 2 sonic anemometers at 29.5 m is facing southeast, while the other faces toward northwest.

2.2 Description of the instrumentation

Parameter	Model	Manufacturer
Air temperature	HMP-45D	Vaisala
Air pressure	PTB-210	Vaisala
Sensible heat flux	DA-600	Kaijo Inc

2.3 Instrumentation specification

Parameter	Sensor Type	Height of sensor (m)	Accuracy	Resolution
Air temperature		1.6, 29.5	0.2 deg	0.1 deg
Air pressure	silicon capacitive sensor	1.5	0.30 hPa	0.1hPa
Sensible heat flux	Sonic Anemometer-thermometer (SAT)	1.6, 29.5	.	1W/m2

3.0 DATA COLLECTION AND PROCESSING

3.1 Description of data collection

Measurements of sonic anemometers are sampled at 10Hz, and those of the other sensors at every 10 seconds, and they are averaged over 30 minutes in the loggers. Averaged data stored in the these loggers are retrieved simultaneously through the local network and they are stored in the data server of TERC. These data are subject to the post-processed for CEOP2

3.2 Description of derived parameters and processing techniques used

The observations from the two sonic anemometers at 29.5 m on the tower are both recorded, and one of the two was selected with the wind direction as a reference during the post-processing. Sensible heat flux was computed from the w-T covariance measured with SAT, air temperature at the same level, and atmospheric pressure at 1.5 m.

4.0 QUALITY CONTROL PROCEDURES

Measurements during the maintenance of the instruments as well as other possible causes such as electric shutdown are marked as “M” (missing), and the data value is replaced with the value -999.99. The measurements that are out of the ordinary range are also marked as “M” (missing). Some abnormal data are marked as “D”(dubious). Otherwise, the data is marked as “G” (good).

5.0 GAP FILLING PROCEDURES

No gap filling procedure was applied.

6.0 DATA REMARKS

6.1.1 Instruments problems

6.1.2 Quality issues

6.2 Missing data periods

Regular electric shutdown occurred between Oct 17 and Oct 20, and no data were recorded during this time.

Grass mowing works on July 15-16, and Oct 25-27.

Regular maintenance was made on Aug 1st, 13rd, Sep 2nd, 10th, 16th, Oct 2nd, 16th, Nov 4th, 18th, Dec 1st, 9th, and 18th.

Quality flags are marked “D”, “M” or “G” depending on the possible influence on the measured values.

7.0 REFERENCE REQUIREMENTS

8.0 REFERENCES