

PCASP-1 (aka, the UWYO/IBR PCASP)

Cai et al. (2013) and Snider et al. (2017) explain how a PCASP is challenged with laboratory-generated monodisperse test particles. This was done before and after the TRANS²AM22 deployment. The Table has sizing derived using polystyrene latex test particles (refractive index = 1.59).

The table has upper-bound particle sizes for 29 channels. The lower bound for channel 1 is 0.09 micrometer. The aerosol sample flow rate calibration did not change significantly before-to-after the TRANS²AM-22 deployment.

Cai, Y., J.R.Snider and P. Wechsler, Calibration of the passive cavity aerosol spectrometer probe for airborne determination of the size distribution, Atmos. Meas. Tech., 6, 2349–2358, 2013

Snider, J.R., D.Leon and Z.Wang, Droplet Concentration and Spectral Broadening in Southeast Pacific Stratocumulus, J. Atmos. Sci., 74, 719-749, 2017

Channel Number	Size Calibration
	Diameter, micrometer
1	0.10
2	0.11
3	0.12
4	0.13
6	0.14
7	0.15
8	0.16
9	0.17
10	0.19
11	0.21
12	0.23
13	0.25
14	0.27
15	0.29
16	0.45
17	0.55
18	0.65
19	0.75
20	0.85
21	0.95
22	1.05
23	1.25
24	1.45
25	1.65
26	1.85
27	2.05
28	2.35
29	2.65
30	3.05

 Channel 5 removed